

# Calibration Cylinder Report

DOT/TC Registration # F616

# Cylinder Sales and Testing, LLC

7120 Golden Ring Road

Baltimore, MD 21221

**Disposition Explanation:** PCAL=passed, FCAL=failed, PAA=aborted

Operator Name: \_\_\_\_\_ Signature: \_\_\_\_\_ Stamp: \_\_\_\_\_ Date Signed: \_\_\_\_\_  
Certifying \_\_\_\_\_ Signature: \_\_\_\_\_ Stamp: \_\_\_\_\_ Date Signed: \_\_\_\_\_  
Staff/Inspector Name: \_\_\_\_\_

I hereby certify that all the following tests were made under my supervision and in accordance with DOT/CTC regulations.

#	Test Date Test Time	Operator	Serial Number	Cyl MFG.	Specified Unit	Actual	Test Time	TE Max TE Min	Total Elastic	Perm Percen	Disposition Remark
1	2/17/2021 8:20		LCC1710-1314B		3000 PSI	3016	30	56.8 55.8	56.6 56.6	0.0 0.0	PCAL Jkt #1
2	2/17/2021 8:22		LCC1710-1314B		4000 PSI	4005	30	75.8 74.4	75.3 75.3	0.0 0.0	PCAL Jkt #1
3	2/17/2021 8:25		LCC1710-1314B		5000 PSI	5013	30	95.1 93.3	94.5 94.5	0.0 0.0	PCAL Jkt #1
4	2/17/2021 8:27		LCC1710-1314B		6000 PSI	6007	30	114.3 112.1	113.4 113.4	0.0 0.0	PCAL Jkt #1

# Cylinder Requalification Report

DOT/TC Registration # F616

## Cylinder Sales and Testing, LLC

7120 Golden Ring Road

Baltimore, MD 21221

**Code Has Three Letters:** 1st Letter: Visual Test, 2nd Letter: % Perm Exp Test, 3rd Letter: REE Test

**Letter Code:** P=Passed, F=Failed, N=Not Applicable, A=Aborted **Example:** PPP-Passed Visual, Passed % Prm, Passed REE

**PCAL, FCAL** used for passed, failed calibration verification test, respectively; **PASS/FAIL** used proof tests-no REE, no % perm criteria

Operator Signature: \_\_\_\_\_ Date Signed: \_\_\_\_\_

Supervisor Signature: \_\_\_\_\_ Date Signed: \_\_\_\_\_

I hereby certify that all the following tests were made under my supervision and in accordance with DOT/CTC regulations.

#	Test Date Test Time	Operator	Serial Number Cyl Owner	Size Service	Cyl MFG. MFG. Date	REE Source	Rating Unit	Specified Actual	Test Time	Total Elastic	Perm Percen	+* VE	VE Date Ref Ring	Disposition Remark
1	2/17/2021 9:11		1A809033 MWS	7.3X21 AIR	PST 07/1980	66.3 CYL	E9791-3442 PSI	5250 5276	30	67.0 59.9	7.1 10.6			PFP condemned
3	2/17/2021 10:01		1A809491 MWS	7.3X21 AIR	PST 12/2004	66.3 CYL	E9791-3442 PSI	5250 5265	30	60.1 59.6	0.5 0.8			PPP
4	2/17/2021 10:58		4462247 MWS	7.3X26 AIR	M8004 09/2010	62 CYL	3AA2400 PSI	4000 4024	30	57.7 57.0	0.7 1.2			PPP
5	2/17/2021 11:09		4462244 MWS	7.3X26 AIR	M8004 09/2010	62 CYL	3AA2400 PSI	4000 4018	30	57.3 56.9	0.4 0.7			PPP